



Important Notes and Recommendations

- Only the natural tooth measure
- Lamp off, avoid intense light sources
- Patient must be seated in an upright position
- Tooth must be moist and clean
- Measuring tube must be flush to the tooth
- Shade Star does not wobble
- Upper lip hang up on the measuring tube



Correct



Flush to surface of tooth

Incorrect



Incorrect positioning of the tooth - instrument tilted away from tooth allowing ambient light into measurement

